



Quantitative MRI (qMRI)

Govind Nair


Staff Scientist, NINDS

Neurodegenerative Changes




Multiple sclerosis is an immune mediated neurodegenerative disease affecting the myelin, axons, and neurons.

Qualitative vs. Quantitative



Periventricular hypointensity on T1 MPRAGE.





Area: 42.809 mm²
Mean: 107.464 SDev: 17.948 Sum: 6018
Min: 79.000 Max: 165.000
Length: 2.351 cm

Area: 2.455 cm²
Mean: 97.255 SDev: 28.968 Sum: 25578
Min: 52.000 Max: 169.000
Length: 6.120 cm

Area: 2.055 cm²
Mean: 112.075 SDev: 22.097 Sum: 25441
Min: 56.000 Max: 170.000
Length: 5.372 cm


The Trouble with Quantitation



N=3 on each scanner

Different scanners, similar protocols
FLAIR

The Trouble with Quantitation



Different scanners, very similar protocols
Normalized signal from FLAIR





Coil Sensitivities Effect Normalization

Receiver effects





Transmit effects



(Images of a ball of water should be uniform)

Coil Sensitivities Effect Normalization

High signal
Near receive coil



Poor signal
Far from transmit/receive coil



Why Bother with Quantitation: Philosophical

"I often say that when you can measure what you are speaking about, and express it in numbers, you know something about it; but when you cannot measure it, when you cannot express it in numbers, your knowledge is of a meager and unsatisfactory kind; it may be the beginning of knowledge, but you have scarcely in your thoughts advanced to the state of Science, whatever the matter may be."

- *Lord Kelvin [PLA, vol. 1, "Electrical Units of Measurement", 1883-05-03]*


(Pre)clinically Available qMRI

qMRI technique	Biological processes affecting them
Diffusion Tensor Imaging (DTI)	Demyelination, axonal loss, vasogenic edema, ischemia, inflammation...
Magnetization Transfer Ratio (MTR/MTC)	Macromolecular composition, cellularity, edema, iron accumulation...
MRI relaxometry (qT ₁ , qT ₂ , qT ₂ [*])	Demyelination, gliosis, tissue loss, iron accumulation, edema, macromolecular composition
Quantitative susceptibility mapping (QSM)	Demyelination, iron accumulation...
Dynamic contrast-enhanced MRI (DCE)	Blood Brain Barrier permeability...
MR Spectroscopy (qMRS)	Neuronal loss (NAA), glial cell activation (ml), lactate accumulation (Lac), cellular debris, infections...
Labeling with MRI contrast agents (Iron oxide, Mn)	Cellular migration or tracking, cellular activation (when conjugated with Ab)...
Volumetrics	Atrophy, segmentation errors, edema, pressure...

Remember: robust, repeatable, and biologically relevant

Quantitative MRI

- Robust, repeatable, and biologically relevant.
- Independent of scanner, software, hardware.



Laboratory results

WBC	4.57	[4.23-9.07 K/uL]
RBC	4.36 ↓	[4.63-6.08 M/uL]
HGB	13.2 ↓	[13.7-17.5 g/dL]
HCT	37.8 ↓	[40.1-51.0 %]
MCV	86.7	[79.0-92.2 fL]
MCH	30.3	[25.7-32.2 pg]
MCHC	34.9	[32.3-36.5 g/dL]
RDW	11.8	[11.6-14.4 %]
Platelet Count	256	[161-347 K/uL]
MPV	10.3	[9.4-12.4 fL]
Nucleated RBC	0.0	[0.0-0.2 /100 WBC]
Nucleated RBC Absolute	0.00	[0.00-0.01 K/uL]
Neutrophils	45.4	[34.0-67.9 %]
Bands		with Neutrophil
Immature Granulocytes	0.2	[0.0-0.4 %]
Lymphocytes	43.5	[21.8-53.1 %]
Monocytes.	8.3	[5.3-12.2 %]
Eosinophils	2.2	[0.8-7.0 %]
Basophils	0.4	[0.2-1.2 %]
Neutrophil Absolute	2.07	[1.78-5.38 K/uL]
Immature Granulocytes Absolute	0.01	[0.00-0.03 K/uL]
Lymphocyte Absolute	1.99	[1.32-3.57 K/uL]
Monocyte Absolute	0.38	[0.30-0.82 K/uL]
Eosinophil Absolute	0.10	[0.04-0.54 K/uL]
Basophil Absolute	0.02	[0.01-0.08 K/uL]

qMRI results

qMRI parameter	Subject	Normative range
Grey matter volume	750 cc	[600-800 cc]*
FA White matter	0.65	[0.5-0.8]*
T1 GM	1523 ms	[1200-1600 ms]*
...		

*made up values

qMRI in Neuroinflammation

Morphometry

- Atrophy of the brain.
- Atrophy of the spinal cord.
- Lesion volume.


Microstructural changes

- Relaxometry (T_1 , T_2 , T_2^*)
- Diffusion Tensor Imaging
- Magnetization Transfer Ratio
- Spectroscopy
- Functional connectivity
- ...

Inflammatory markers



- Blood perfusion imaging
- BBB permeability

MRI Basics




Quick Review of Basic MRI Contrasts

Main Magnetic Field





Quick Review of Basic MRI Contrasts




Biological changes are likely to change relaxation properties.

Measuring Rate of T_1 Relaxation

Inversion recovery (IR)






$$M = M_0 * (1 - 2 \alpha * e^{-TI/T1})$$



Inversion Pulse

Inversion Time (TI) from 150 ms to 450 ms




T₁ map

$$M = M_0 * (1 - 2\alpha * e^{-TI/T1})$$

M₀ also has receiver gain effects
α is transmit effects

Measuring Rate of T_1 Relaxation



$$S = M_0 \frac{(1 - e^{-TR/T_1}) \sin \theta}{1 - e^{-TR/T_1} \cos \theta}$$



θ is the flip angle and S the signal at that flip

$$\frac{M_\theta}{\sin \theta} = e^{-T/T_1} \frac{M_\theta}{\tan \theta} + M_0(1 - e^{-T/T_1})$$

Of the form: $Y = bX + a$

$$T_1 = -\frac{TR}{\ln b}$$

However, transmit coil profiles are not corrected automatically since FA needs to be specified.



Line fitting


Correcting for B1

Uncorrected T1 map



0.5s

0



5s


0

Corrected T1 map



0.5s

0



5s


0

MP2RAGE Morphometry


Raw image, inv1



Raw image, inv2



T1 map



Segmentation



Brain Morphometry Report - 1/6

WIP900B_VD13A.30

Patient Demographics	32 yrs	Male	
Image quality	high	0.70	[0 - 0.82]
Segmentation Quality	high	0.74	[0.7 - 1]
Tissue	Absolute [ml]	Normalized^ [%]	Normative Range
TIV	1477.6		
GM	654.3	* 44.3	[45.6 - 53.2]
cortical GM	470.6	* 31.8	[34.6 - 41.2]
WM	406.0	* 27.5	[28.1 - 33.9]
WMab	24.1	1.6	
CSF	417.2	* 28.2	[17.1 - 22.7]

^ Percentage of TIV (Total Intracranial Volume)

^^ 10th and 90th percentiles of healthy age-matched population

* Out-of-range volumes

WARNING MP2RAGE UNI-DEN... see log file for more details.

not approved for diagnostic purpose

Brain Morphometry Report - 3/6

Structure	Absolute [ml]	Normalized^ [%]	Normative Range^^
Hippocampus	6.5	0.44	[0.40 - 0.51]
Hippocampus left	3.3	0.22	[0.20 - 0.26]
Hippocampus right	3.2	0.22	[0.20 - 0.26]
Ventricles	60.4	* 4.09	[0.88 - 2.45]
lateral ventricle left	30.6	* 2.07	[0.35 - 1.09]
lateral ventricle right	22.8	* 1.54	[0.33 - 1.01]
3rd ventricle	3.3	* 0.22	[0.07 - 0.14]
4th ventricle	3.7	* 0.25	[0.11 - 0.22]




not approved for diagnostic purpose

Brain Morphometry Report - 2/6

Structure	Absolute [ml]	Normalized^ [%]	Normative Range^^ [%]
Thalamus	16.2	1.09	[0.92 - 1.11]
Thalamus left	7.7	0.52	[0.45 - 0.55]
Thalamus right	8.5	0.57	[0.46 - 0.57]
Putamen	13.3	0.90	[0.88 - 1.10]
Putamen left	6.8	0.46	[0.46 - 0.57]
Putamen right	6.5	0.44	[0.42 - 0.53]
Caudate	9.3	0.63	[0.54 - 0.71]
Caudate left	4.7	0.32	[0.26 - 0.35]
Caudate right	4.6	0.31	[0.28 - 0.37]
Pallidum	3.7	0.25	[0.25 - 0.32]
Pallidum left	1.7	0.12	[0.12 - 0.16]
Pallidum right	2.0	0.14	[0.12 - 0.16]
Deep WM left	11.4	0.77	[0.66 - 0.83]
Deep WM right	9.9	0.67	[0.57 - 0.71]

not approved for diagnostic purpose


Transverse (T_2) Relaxation





Free Induction Decay

Generating an Echo


Spin Echo



Gradient Echo





T2 vs. T2*





Signal loss due to:

- (Macroscopic) magnetic field inhomogeneities (refocused by the 180° pulse)
- Local environment (presence of paramagnetic molecules, viscosity...) – T₂


Microscopic and Macroscopic Effects




Measuring Rate of T_2/T_2^* Relaxation




Measuring Rate of T_2^* Relaxation




TE= 6 ms




16 ms



26 ms





36 ms





T_2 map

MR Fingerprinting (Also MAGiC*)




Multiparametric Approach to Improve Specificity



GBM with vasogenic edema






Typical distribution




Suspected Leukoencephalopathy



Biological Relevance




Multi-compartment model



1 mm

Water Exchange Through Compartments



qMRI in Neuroinflammation

Morphometry

- Atrophy of the brain.
- Atrophy of the spinal cord.
- Lesion volume.


Microstructural changes

- Relaxometry (T_1 , T_2 , T_2^*)
- Diffusion Tensor Imaging
- Magnetization Transfer Ratio
- Spectroscopy
- Functional connectivity
- ...

Inflammatory markers

- Blood perfusion imaging
- BBB permeability

Cerebral Atrophy in Multiple Sclerosis



Female
Clinically diagnosed
with multiple
sclerosis
YOB: 1963

1995

2014

MS patient – T2-weighted images, 19 years apart



“[Atrophy] is the ultimate consequence of destructive pathological changes... within lesions or in normal appearing tissue”: Miller et al Brain (2002) 125: 1677

Methods for Measuring Brain Morphologic Features on Magnetic Resonance Images


Validation and Normal Aging

Terry L. Jernigan, PhD; Gary A. Press, MD; John R. Hesselink, MD

(*Arch Neurol.* 1990;47:27-32)



Cerebral Proportions	Operator 1, Mean ± SD	Operator 2, Mean ± SD	Spearman Rank Order
Fluid	0.11 ± 0.06	0.10 ± 0.06	.98
Gray Matter	0.52 ± 0.06	0.52 ± 0.08	.92
White matter	0.37 ± 0.05	0.38 ± 0.07	.84
Signal hyperintensity	0.002 ± 0.001	0.002 ± 0.001	.86



Imaging protocol

Exclusion of non-brain areas

Pixel classification

Calculate volume, reliability analysis

Table 1. Methods Used for Whole-Brain Atrophy Measurement in Multiple Sclerosis


Method	Segmentation	Registration	Normalization	Automation	Comments
Brain parenchymal fraction	Brain parenchyma, ventricular CSF	No	Brain + ventricular CSF	Full	Used on commonly acquired MR images Includes only ventricular CSF
Index of brain atrophy	Brain parenchyma, ventricular CSF	No	Brain + ventricular and sulcal CSF	Semi	Only measures above midbrain High-resolution images
Whole-brain ratio	Intradural volume, CSF volume	No	Intradural volume	Semi	Manual editing of lesions
Brain to intracranial capacity ratio	Gray matter, white matter, lesions, CSF; Bayesian tissue classification	Yes	Intracranial volume	Full	Limited coverage in reported cases Intensity correction
3DVIEWNIX	Gray matter, white matter, lesions, CSF; fuzzy connectedness-based thresholding	Yes	Intracranial volume	Semi	Time-consuming operator input Intensity correction
Statistical parametric mapping	Gray matter, white matter, CSF; stereotactic space	Yes	Intracranial volume	Semi for MS lesions	Manual editing of misclassified voxels
Template-driven segmentation	Template-driven, brain parenchyma, CSF	Yes	Intracranial volume	Full	Limited application in MS
Alfano	Gray matter, white matter, lesions, CSF; relaxometric characterization	No	Intracranial volume	Full	Intensity correction
Structural image evaluation using normalization of atrophy X/SIENA	Brain and skull	Yes	Head size	Full	No CSF segmentation needed
Brain boundary shift integral	No	Yes	Brain size	Semi and full versions	Strongly depends on accuracy of registration No segmentation needed
Voxel-based morphometry	Gray matter, white matter, CSF	Yes	Intracranial volume (possible)	Full	Lesion mask needed for white matter analysis Complex statistical analysis

CSF = cerebrospinal fluid, MR = magnetic resonance, MS = multiple sclerosis.


Need for automation, especially in large cohort studies

Couple of Examples


BPF



SIENAX




Volumetrics - LesionTOADS



MPRAGE




2D FLAIR




Tissue classification

Global Cerebral Atrophy – Brain Free Water Imaging




FLAIR – unprocessed
Generally, 1 mm isotropic




BFWI - unprocessed
The only thing that is bright is fluids
Done at 0.65 mm isotropic

Comparison: BFWI vs. LesionTOADS



Original




LesionTOADS - processed



BFWI - processed

What does it mean clinically?



EDSS: Kurtzke Expanded Disability Status Scale

SNRS: Scripps Neurologic Rating Scale

SDMT: Symbol Digit Modalities Test

PASAT: Paced Auditory Serial Addition Test


9HPT: 9-Hole Peg Test

25' walk: 25-foot Walk Test

Atrophy of the Spinal Cord



38 y.o. male, healthy volunteer






31 y.o. female with MS



Atrophy in MS

In comparison - 38% smaller cross-sectional area






HAM/TSP (n=18)	C1	C2	C3	C4	C5	C6	C7	T1	T2	T3	T4	T5	T6	T7	T8	T9	T10	C	T
SNRS	0.14	0.14	0.22	0.24	0.09	0.22	0.16	0.29	0.20	0.14	0.22	0.18	0.25	0.29	0.19	-0.03	0.01	0.20	0.19
EDSS	-0.33	-0.36	-0.39	-0.46	-0.44	-0.44	-0.46	-0.56	-0.57	-0.55	-0.51	-0.50	-0.59	-0.50	-0.36	-0.22	-0.15	-0.45	-0.48
IPEC	0.07	0.06	-0.05	-0.09	-0.01	-0.01	0.07	0.02	-0.07	0.02	-0.04	-0.02	-0.02	0.02	0.03	0.05	0.06	0.00	0.00
Ambulation Index	-0.46	-0.53	-0.54	-0.58	-0.58	-0.46	-0.46	-0.57	-0.65	-0.57	-0.59	-0.61	-0.72	-0.63	-0.49	-0.44	-0.34	-0.54	-0.57
Disease Duration	-0.66	-0.62	-0.63	-0.64	-0.61	-0.65	-0.61	-0.62	-0.58	-0.52	-0.51	-0.48	-0.47	-0.46	-0.32	-0.30	-0.32	-0.68	-0.49
Proviral load	-0.02	-0.14	-0.13	-0.14	-0.09	-0.24	-0.27	-0.30	-0.23	0.01	-0.19	-0.18	-0.04	-0.06	0.07	-0.01	-0.18	-0.17	-0.13
MS (n=18)																			
SNRS	0.62	0.65	0.61	0.58	0.52	0.56	0.49	0.66	0.61	0.61	0.71	0.70	0.63	0.65	0.53	0.49	0.23	0.63	0.64
EDSS	-0.75	-0.67	-0.63	-0.57	-0.51	-0.51	-0.44	-0.53	-0.54	-0.54	-0.52	-0.58	-0.58	-0.51	-0.45	-0.43	-0.28	-0.61	-0.55


HAM/TSP: Human T-cell lymphotropic virus type 1 Associated Myelopathy/Tropical Spastic Paraparesis

Longitudinal Monitoring of Cord Atrophy


2/27/13




4/24/13, AI=4




3/6/14, AI=5



6/4/14



3/19/15



37 y.o. Jamaican female with clinical diagnosis of HAM/TSP.

Symptom start 8/2012, progressive weakening.

Cervical: 7.8% reduction/year.
Thoracic: 11% reduction/year.

Summary

- Several relevant qMRI measurements are readily available on most modern scanners.
 - Important to understand the imaging protocol and analysis methods for reliable measurement.
- Some qMRI measures are more specific to biological processes than others.
 - Multiparametric techniques may offer more specificity and a better understanding of the biological processes.
- Longitudinal measurements may be more fruitful.

Thank you.